Non responsive based on revised scope

Certificate of Analysis

Final Report

Laboratory Order ID 19J0229

Client Name: Non responsive based on revised scope

Submitted To:

Client Site I D Shiloh Church Rd Site

Date Received: October 4, 2019 15:00

Date I ued October 9, 2019 16:43

Project Number: [none]

Purchase Order: 64073

Enclosed are the results of analyses for samples received by the laboratory on 10/04/2019 15:00. If you have any que tion concerning thi report, plea e feel free to contact the laboratory

Sincerely,

Non responsive based on revised scope

Technical Director

End Notes:

The test results listed in this report relate only to the samples submitted to the laboratory and as received by the Laboratory.

Unless otherwise noted, the test results for solid materials are calculated on a wet weight basis. Analyses for pH, dissolved oxygen, temperature, residual chlorine and sulfite that are performed in the laboratory do not meet NELAC requirements due to extremely short holding times. These analyses should be performed in the field. The results of field analyses performed by the Sampler included in the Certificate of Analysis are done so at the client's request and are not included in the laboratory's fields of certification nor have they been audited for adherence to a reference method or procedure.

The signature on the final report certifies that these results conform to all applicable NELAC standards unless otherwise specified. For a complete list of the Laboratory's NELAC certified parameters please contact customer service.

This report shall not be reproduced except in full without the expressed and written approval of an authorized representative of Non R
Non responsive based on revised scope

ACCREDITED IN ACCORDANCE WITH nelac Commonwealth of Virginia, DCLS Accreditation # 460021

NWBOC EDWOSB CERTIFIED

CERTIFIED SWaM Small, Women and Minority-Owned Supplier Diversity Strengthens the Commonwealth by the Virginia Department of Minority Business Enterprise

Certificate of Analysis

Final Report

Client Name	Non responsive based on revised scope	Date I ued	10/9/2019 16:43
Submitted To		Project Number	[none]
Client Site I.D.:	Shiloh Church Rd Site	Purcha e Order	64073

ANALYTICAL REPORT FOR SAMPLES

Laboratory Order ID 19J0229

Sample ID	Laboratory ID	Matrix	Date Sampled	Date Received
W-Grab	19J0229-01	Soil	10/04/2019 12:00	10/04/2019 15:00
E-Grab	19J0229-02	Soil	10/04/2019 12:00	10/04/2019 15:00

PCB results have been calculated based on dry weight.

[redacted]

Certificate of Analysis

Final Report

Client Name: [redacted]

Submitted To:

Client Site I.D.: Shiloh Church Rd Site

Date Issued: 10/9/2019 16:43

Project Number: [none]

Purchase Order: 64073

Laboratory Order ID: 19J0229

Analytical Results

Sample I.D. W-Grab **Laboratory Sample ID:** 19J0229-01

Grab Date/Time 10/04/2019 12:00

Field Residual Cl: **Field pH:**

Parameter	Samp ID	Method	Result	Qual	Reporting Limit	D.F.	Sample Prep Date/Time	Analysis Date/Time	Analyst
Organochlorine Pesticides and PCBs by GC/ECD									[redacted]
PCB as Aroclor 1016	01	SW8082A	<0.556 mg/kg dry		0.556	5	10/07/19 09:45	10/07/19 21:57	
PCB as Aroclor 1221	01	SW8082A	<0.556 mg/kg dry		0.556	5	10/07/19 09:45	10/07/19 21:57	
PCB as Aroclor 1232	01	SW8082A	<0.556 mg/kg dry		0.556	5	10/07/19 09:45	10/07/19 21:57	
PCB as Aroclor 1242	01	SW8082A	<0.556 mg/kg dry		0.556	5	10/07/19 09:45	10/07/19 21:57	
PCB as Aroclor 1248	01	SW8082A	<0.556 mg/kg dry		0.556	5	10/07/19 09:45	10/07/19 21:57	
PCB as Aroclor 1254	01RE1	SW8082A	**152 mg/kg dry**		55.6	500	10/07/19 09:45	10/09/19 11:34	
PCB as Aroclor 1260	01	SW8082A	<0.556 mg/kg dry		0.556	5	10/07/19 09:45	10/07/19 21:57	
Surr: DCB	*01*	*SW8082A*	*200 %*	*DS*	*30-105*		*10/07/19 09:45*	*10/07/19 21:57*	
Surr: TCMX	*01*	*SW8082A*	*75.0 %*		*30-105*		*10/07/19 09:45*	*10/07/19 21:57*	
Wet Chemistry Analysis									
Percent Solids	01	SM18 2540G	**89.3 %**		0.10	1	10/08/19 14:20	10/08/19 14:20	

[redacted]

Certificate of Analysis

Final Report

Client Name: [redacted]

Submitted To:

Client Site I.D.: Shiloh Church Rd Site

Date Issued: 10/9/2019 16:43

Project Number: [none]

Purchase Order: 64073

Laboratory Order ID: 19J0229

Analytical Results

Sample I.D. E-Grab

Laboratory Sample ID: 19J0229-02

Grab Date/Time 10/04/2019 12:00

Field Residual Cl:

Field pH:

Parameter	Samp ID	Method	Result	Qual	Reporting Limit	D.F.	Sample Prep Date/Time	Analysis Date/Time	Analyst
Organochlorine Pesticides and PCBs by GC/ECD									[redacted]
PCB as Aroclor 1016	02	SW8082A	<0.559 mg/kg dry		0.559	5	10/07/19 09:45	10/07/19 22:16	
PCB as Aroclor 1221	02	SW8082A	<0.559 mg/kg dry		0.559	5	10/07/19 09:45	10/07/19 22:16	
PCB as Aroclor 1232	02	SW8082A	<0.559 mg/kg dry		0.559	5	10/07/19 09:45	10/07/19 22:16	
PCB as Aroclor 1242	02	SW8082A	<0.559 mg/kg dry		0.559	5	10/07/19 09:45	10/07/19 22:16	
PCB as Aroclor 1248	02	SW8082A	<0.559 mg/kg dry		0.559	5	10/07/19 09:45	10/07/19 22:16	
PCB as Aroclor 1254	02RE1	SW8082A	**324 mg/kg dry**		112	1000	10/07/19 09:45	10/09/19 12:59	
PCB as Aroclor 1260	02	SW8082A	<0.559 mg/kg dry		0.559	5	10/07/19 09:45	10/07/19 22:16	
Surr: DCB	*02*	*SW8082A*	*200 %*	*DS*	*30-105*		*10/07/19 09:45*	*10/07/19 22:16*	
Surr: TCMX	*02*	*SW8082A*	*100 %*		*30-105*		*10/07/19 09:45*	*10/07/19 22:16*	
Wet Chemistry Analysis									
Percent Solids	02	SM18 2540G	**88.9 %**		0.10	1	10/08/19 14:20	10/08/19 14:20	

Certificate of Analysis

Final Report

Client Name: Non responsive based on revised scope

Submitted To:

Client Site I.D.: Shiloh Church Rd Site

Date Issued: 10/9/2019 16:43

Project Number: [none]

Purchase Order: 64073

Analytical Summary

Sample ID	Preparation Factors Initial / Final	Method	Batch ID	Sequence ID	Calibration ID
Wet Chemistry Analysis		**Preparation Method:**	**No Prep Wet Chem**		
19J0229-01	10.0 g / 10.0 mL	SM18 2540G	BCJ0276	SCJ0247	
19J0229-02	10.0 g / 10.0 mL	SM18 2540G	BCJ0276	SCJ0247	

Sample ID	Preparation Factors Initial / Final	Method	Batch ID	Sequence ID	Calibration ID
Organochlorine Pesticides and PCBs by GC/ECD		**Preparation Method:**	**SW3550B**		
19J0229-01	30.2 g / 5.00 mL	SW8082A	BCJ0217	SCJ0239	AI90153
19J0229-01RE1	30.2 g / 5.00 mL	SW8082A	BCJ0217	SCJ0285	AI90099
19J0229-02	30.2 g / 5.00 mL	SW8082A	BCJ0217	SCJ0239	AI90153
19J0229-02RE1	30.2 g / 5.00 mL	SW8082A	BCJ0217	SCJ0285	AI90099

Certificate of Analysis

Final Report

Client Name:

Date Issued: 10/9/2019 16:43

Submitted To:

Client Site I.D.: Shiloh Church Rd Site

Project Number: [none]

Purchase Order: 64073

Organochlorine Pesticides and PCBs by GC/ECD - Quality Control

Air Water & Soil Laboratories, Inc.

Analyte	Result	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Qual
Batch BCJ0217 - SW3550B										
Blank (BCJ0217-BLK1)				Prepared & Analyzed: 10/07/2019						
PCB as Aroclor 1016	<0.100 mg/kg wet	0.100	mg/kg wet							
PCB as Aroclor 1221	<0.100 mg/kg wet	0.100	mg/kg wet							
PCB as Aroclor 1232	<0.100 mg/kg wet	0.100	mg/kg wet							
PCB as Aroclor 1242	<0.100 mg/kg wet	0.100	mg/kg wet							
PCB as Aroclor 1248	<0.100 mg/kg wet	0.100	mg/kg wet							
PCB as Aroclor 1254	<0.100 mg/kg wet	0.100	mg/kg wet							
PCB as Aroclor 1260	<0.100 mg/kg wet	0.100	mg/kg wet							
Surr: DCB	0.0415		mg/kg wet	0.0332		125	30-105			S
Surr: TCMX	0.0233		mg/kg wet	0.0332		70.0	30-105			
LCS (BCJ0217-BS2)				Prepared & Analyzed: 10/07/2019						
PCB as Aroclor 1016	0.146 mg/kg wet	0.100	mg/kg wet	0.164	mg/kg wet	89.0	60-140			
PCB as Aroclor 1260	0.167 mg/kg wet	0.100	mg/kg wet	0.164	mg/kg wet	102	60-140			
Surr: DCB	0.0393		mg/kg wet	0.0328	mg/kg wet	120	30-105			S
Surr: TCMX	0.0213		mg/kg wet	0.0328	mg/kg wet	65.0	30-105			
Matrix Spike (BCJ0217-MS2)		**Source: 19J0184-02**		Prepared & Analyzed: 10/07/2019						
PCB as Aroclor 1016	0.189 mg/kg dry	0.111	mg/kg dry	0.185	<0.111 mg/kg dry	102	60-140			
PCB as Aroclor 1260	0.202 mg/kg dry	0.111	mg/kg dry	0.185	<0.111 mg/kg dry	109	60-140			
Surr: DCB	0.0426		mg/kg dry	0.0371	mg/kg dry	115	30-105			S
Surr: TCMX	0.0315		mg/kg dry	0.0371	mg/kg dry	85.0	30-105			
Matrix Spike Dup (BCJ0217-MSD2)		**Source: 19J0184-02**		Prepared & Analyzed: 10/07/2019						
PCB as Aroclor 1016	0.189 mg/kg dry	0.112	mg/kg dry	0.187	<0.112 mg/kg dry	101	60-140	0.00322	20	
PCB as Aroclor 1260	0.208 mg/kg dry	0.112	mg/kg dry	0.187	<0.112 mg/kg dry	111	60-140	2.80	20	
Surr: DCB	0.0431		mg/kg dry	0.0374	mg/kg dry	115	30-105			S
Surr: TCMX	0.0300		mg/kg dry	0.0374	mg/kg dry	80.0	30-105			

Certificate of Analysis

Final Report

Client Name:

Submitted To:

Client Site I.D.: Shiloh Church Rd Site

Date Issued: 10/9/2019 16:43

Project Number: [none]

Purchase Order: 64073

Wet Chemistry Analysis - Quality Control

Air Water & Soil Laboratories, Inc.

Analyte	Result	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Qual
Batch BCJ0276 - No Prep Wet Chem										
Blank (BCJ0276-BLK1)				Prepared & Analyzed 10/08/2019						
Percent Solids	100 %	0.10	%							
Duplicate (BCJ0276-DUP1)		**Source: 19J0229-02**		Prepared & Analyzed: 10/08/2019						
Percent Solids	88.5 %	0.10	%		88.9 %			0.385	20	
Duplicate (BCJ0276-DUP2)		**Source: 19J0276-01**		Prepared & Analyzed: 10/08/2019						
Percent Solids	86.1 %	0.10	%		84.6 %			1.74	20	



Certificate of Analysis

Final Report

Client Name:		Date Issued:	10/9/2019 16:43
Submitted To:		Project Number:	[none]
Client Site I.D.:	Shiloh Church Rd Site	Purchase Order:	64073

Certified Analyses included in this Report

Analyte	Certifications
SW8082A in Solids	
PCB as Aroclor 1016	VELAP,NC
PCB as Aroclor 1221	VELAP,NC
PCB as Aroclor 1232	VELAP,NC
PCB as Aroclor 1242	VELAP,NC
PCB as Aroclor 1248	VELAP,NC
PCB as Aroclor 1254	VELAP,NC
PCB as Aroclor 1260	VELAP,NC

Code	Description	Lab Number	Expires
MdDOE	Maryland DE Drinking Water	341	12/31/2019
NC	North Carolina DENR	495	12/31/2019
VELAP	NELAC-Virginia Certificate #10503	460021	06/14/2020
WVDEP	West Virginia DEP	350	11/30/2019

Non responsive based on revised scope

Certificate of Analysis

Final Report

Client Name:	Non responsive based on revised scope	Date Issued:	10/9/2019 16:43
Submitted To:		Project Number:	[none]
Client Site I.D.:	Shiloh Church Rd Site	Purchase Order:	64073

Summary of Data Qualifiers

DS	Surrogate concentration reflects a dilution factor.
S	Surrogate recovery was outside acceptance criteria
RPD	Relative Percent Difference
Qual	Qualifers
-RE	Denotes sample was re-analyzed
D.F.	Dilution Factor. Please also see the Preparation Factor in the Analysis Summary section.
TIC	Tentatively Identified Compounds are compounds that are identified by comparing the analyte mass spectral pattern with the NIST spectral library . A TIC spectral match is reported when the pattern is at least 75% consistent with the published pattern. Compound concentrations are estimated and are calculated using an internal standard response factor of 1.
PCBs, Total	Total PCBs are defined as the sum of detected Aroclors 1016, 1221, 1232, 1248, 1254, 1260, 1262, and 1268.

Non responsive based on revised scope

Non responsive based on revised scope

Chain of Custody
Effective: Nov 15, 2018

CHAIN OF CUSTODY

COMPANY	Non responsive based on revised scope	INVOICE TO	Non responsive based on revised scope
CONTACT		INVOICE CO	
ADDRESS		INVOICE AD	
PHONE #:		INVOICE PH	
FAX #:	EMAIL	Non responsive base	

PROJECT NAME/Quote #:
SITE NAME: Shiloh Church Rd Site
PROJECT NUMBER:
P.O. #: 64073
Pretreatment Program:

Is sample for compliance reporting? YES (NO) | Regulatory State: | Is sample from a chlorinated supply? YES NO | PWS I.D. #:

SAMPLER NAME (PRINT): Non responsive based on revised scope | SAMPLER SIGNATURE: Non responsive based on revised scope | Turn Around Time: Circle 10 5 Days or 3 Day(s)

Matrix Codes: WW=Waste Water/Storm Water GW=Ground Water DW=Drinking Water S=Soil/Solids OR=Organic A=Air WP=Wipe OT=Other_____

CLIENT SAMPLE I.D.	Grab	Composite	Field Filtered (Dissolved Metals)	Composite Start Date	Composite Start Time	Grab Date or Composite Stop Date	Grab Time or Composite Stop Time	Time Preserved	Matrix (See Codes)	Number of Containers	ANALYSIS / (PRESERVATIVE) PCB (Grab)							
1) W-Grab	X					10/4/19	12:00		S	2	X							
2) E-Grab	X					1	1		S	1	X							
3)																		
4)																		
5)																		
6)																		
7)																		
8)																		
9)																		
10)																		

COMMENTS

Preservative Codes: N=Nitric Acid C=Hydrochloric Acid S=Sulfuric Acid H=Sodium Hydroxide A=Ascorbic Acid Z=Zinc Acetate T=Sodium Thiosulfate M=Methanol

PLEASE NOTE PRESERVATIVE(S), INTERFERENCE CHECKS or PUMP RATE (L/min)

Non responsive based on revised scope	DATE / TIME 10/4/19 1500	Non responsive based on revised scope	DATE / TIME 10/4/19 1500
RELINQUISHED:	DATE / TIME	RECEIVED:	DATE / TIME
RELINQUISHED:	DATE / TIME	RECEIVED:	DATE / TIME

QC Data Package
Level III ☐
Level IV ☐

LAB USE ONLY Therm ID: 371 COOLER TEMP 1.3 °C
Custody Seals used and intact? (Y / N) Received on ice? (Y / N)

Non responsive based on revised scope

Recd: 10/04/2019 Due: 10/09/2019

v130325002 v181115.xls

Page 10 of 11

Certificate of Analysis

Final Report

Client Name:

Date Issued: 10/9/2019 16:43

Submitted To:

Client Site I.D.: Shiloh Church Rd Site

Project Number: [none]

Purchase Order: 64073

Sample Conditions Checklist

Samples Received at:	**1.30°C**
How were samples received?	Walk In
Were Custody Seals used? If so, were they received intact?	No
Are the custody papers filled out completely and correctly?	Yes
Do all bottle labels agree with custody papers?	Yes
Is the temperature blank or representative sample within acceptable limits or received on ice, and recently taken?	Yes
Are all samples within holding time for requested laboratory tests?	Yes
Is a sufficient amount of sample provided to perform the tests included?	Yes
Are all samples in appropriate containers for the analyses requested?	Yes
Were volatile organic containers received?	No
Are all volatile organic and TOX containers free of headspace?	NA
Is a trip blank provided for each VOC sample set? VOC sample sets include EPA8011, EPA504, EPA8260, EPA624, EPA8015 GRO, EPA8021, EPA524, and RSK-175.	NA
Are all samples received appropriately preserved? Note that metals containers do not require field preservation but lab preservation may delay analysis.	Yes